

QUALITY CONTROL CERTIFICATE

PRODUCT NAME: MYCOVIEW SCREENING QUANTUM CODE: 2060Q

LOT NUMBER: 20180201A2SC EXPIRY DATE: 07/09/2019

	LOT N°	EXPIRY DATE (dd/mm/yyyy)
T Broth	20180115	14/10/2019
Mycoview Screening Quantum	20171208	07/09/2019
Rayon Swabs	MFG: 171004	03/10/2019

A. PHYSICAL INSPECTION

KIT COMPONENT	CHECKED
T Broth	
Mycoview Screening Quantum	\square
Rayon Swabs	\square

B. INHIBITION TESTS

MEDIUM / Bacteria	C. albicans	E. coli	S. aureus	M. pneumonie
T Broth (48hrs / 37°C)	-	-	-	NA
Mycoview Screening Quantum (48hrs / 37°C) (Reconstituted with T Broth)	-	-	-	-

^{- :} No growth (no colour change and no turbidity appearance)

C. MYCOVIEW SCREENING QUANTUM FERTILITY AND TRAY CONTROL

Ureaplasma urealyticum

Sample	Results	Wells					
	(24 H)	С	Uu	Mh	L	Е	
Uu66 (10 ⁴ -10 ⁵ CCU/ml)	Expected	+	+	-	+	-	
	Obtained	+	+	-	+	-	
Uu66 (10 ² -10 ³ CCU/ml)	Expected	±	-	-	±	-	
	Obtained	±	-	-	±	-	
Uu S8	Expected	+	+	-	+	-	

1

(10 ⁴ -10 ⁵ CCU/mI)	Obtained	+	+	-	+	-
Uu S8 (10 ² -10 ³ CCU/ml)	Expected	±	ı	ı	±	ı
	Obtained	±	1	-	±	
Uu CHB (10 ⁴ -10 ⁵ CCU/ml)	Expected	+	+	-	+	-
	Obtained	+	+	-	+	-
Uu CHB (10 ² -10 ³ CCU/ml)	Expected	±	-	-	±	-
	Obtained	±	-	-	±	-

Mycoplasma hominis

Sample	Results	Wells					
	(24-48 H)	С	Uu	Mh	L	Е	
Mh PG21 (10 ⁴ -10 ⁵ CCU/ml)	Expected	+	-	+	-	+	
	Obtained	+	-	+	-	+	
Mh PG21 (10 ² -10 ³ CCU/ml)	Expected	±	-	-	-	±	
	Obtained	±	-	-	-	±	
Mh CHB (10 ⁴ -10 ⁵ CCU/ml)	Expected	+	-	+	-	+	
	Obtained	+	-	+	1	+	
Mh CHB (10 ² -10 ³ CCU/ml)	Expected	±	-	-	-	±	
	Obtained	±	-	-	-	±	

D. CONCLUSION

We herewith certify that the product has been checked in the Quality Control Department of ZEAKON Diagnostics SARL and has met the established Quality Control acceptance criteria.

Approved by :
Pierre-Olivier WEICHLEIN Date:09/02/2018