

QUALITY CONTROL CERTIFICATE

PRODUCT NAME: DEHYDRATED MYCOVIEW MEDIUM **CODE:** 2060

LOT NUMBER: 20180201A1SC

EXPIRY DATE: 23/08/2019

	LOT N°	EXPIRY DATE (dd/mm/yyyy)
T Broth	20171124	23/08/2019
C Medium	20180117	16/10/2019
Test Gallery	/	/
Paraffin Oil	/	/
Rayon Swabs	MFG:171004	03/10/2019

pH of complete Medium (T Broth + C Medium) = 6.02

A. PHYSICAL INSPECTION

KIT COMPONENT	CHECKED
T Broth	<input checked="" type="checkbox"/>
C Medium	<input checked="" type="checkbox"/>
Test Gallery	<input type="checkbox"/>
Paraffin Oil	<input type="checkbox"/>
Rayon Swabs	<input checked="" type="checkbox"/>
Incubation rack with the lid	<input type="checkbox"/>

B. INHIBITION TESTS

MEDIUM / Bacteria	C. albicans	E. coli	S. aureus	M. pneumoniae
T Broth (48hrs / 37°C)	-	-	-	NA
C Medium (48hrs / 37°C) (Reconstituted with T Broth)	-	-	-	-

- : No growth (no colour change and no turbidity appearance)

C. C MEDIUM FERTILITY AND TRAY CONTROL

Ureaplasma urealyticum

Sample	Results (24 H)	Wells											
		C	Uu	Mh	L	E	ROX	AZM	JM	MNO	DO	OFX	CIP
Uu66 (10 ⁴ -10 ⁵ CCU/ml)	Expected	+	+	-	+	-	-	-	-	+	+	-	+
	Obtained	+	+	-	+	-	-	-	-	+	+	-	+
Uu66 (10 ² -10 ³ CCU/ml)	Expected	±	-	-	±	-	-	-	-	-	-	-	-
	Obtained	±	-	-	±	-	-	-	-	-	-	-	-
Uu S8 (10 ⁴ -10 ⁵ CCU/ml)	Expected	+	+	-	+	-	-	-	-	-	-	-	+
	Obtained	+	+	-	+	-	-	-	-	-	-	-	+
Uu S8 (10 ² -10 ³ CCU/ml)	Expected	±	-	-	±	-	-	-	-	-	-	-	-
	Obtained	±	-	-	±	-	-	-	-	-	-	-	-
Uu CHB (10 ⁴ -10 ⁵ CCU/ml)	Expected	+	+	-	+	-	-	-	-	-	-	-	+
	Obtained	+	+	-	+	-	-	-	-	-	-	-	+
Uu CHB (10 ² -10 ³ CCU/ml)	Expected	±	-	-	±	-	-	-	-	-	-	-	-
	Obtained	±	-	-	-	-	-	-	-	-	-	-	-

Mycoplasma hominis

Sample	Results (24-48 H)	Wells											
		C	Uu	Mh	L	E	ROX	AZM	JM	MNO	DO	OFX	CIP
Mh PG21 (10 ⁴ -10 ⁵ CCU/ml)	Expected	+	-	+	-	+	+	+	-	-	-	-	-
	Obtained	+	-	+	-	+	+	+	-	-	-	-	-
Mh PG21 (10 ² -10 ³ CCU/ml)	Expected	±	-	-	-	±	-	-	-	-	-	-	-
	Obtained	-	-	-	-	-	-	-	-	-	-	-	-
Mh CHB (10 ⁴ -10 ⁵ CCU/ml)	Expected	+	-	+	-	+	+	+	-	-	-	-	-
	Obtained	+	-	+	-	+	+	+	-	-	-	-	-
Mh CHB (10 ² -10 ³ CCU/ml)	Expected	±	-	-	-	±	-	-	-	-	-	-	-
	Obtained	±	-	-	-	±	-	-	-	-	-	-	-

D. CONCLUSION

We herewith certify that the product has been checked in the Quality Control Department of ZEAKON Diagnostics SARL and has met the established Quality Control acceptance criteria.

Approved by :

Pierre-Olivier WEICHLEIN

Date :09/02/2018