

QUALITY CONTROL CERTIFICATE

PRODUCT NAME: MYCOVIEW SCREENING QUANTUM CODE: 2060Q

LOT NUMBER: 20180301A2SC EXPIRY DATE: 07/09/2019

	LOT N°	EXPIRY DATE (dd/mm/yyyy)
T Broth	20171208	07/09/2019
Mycoview Screening Quantum	20180224	23/11/2019
Rayon Swabs	MFG: 171212	11/12/2019

A. PHYSICAL INSPECTION

KIT COMPONENT	CHECKED
T Broth	abla
Mycoview Screening Quantum	$oxed{oxed}$
Rayon Swabs	Image: section of the content of the

B. INHIBITION TESTS

MEDIUM / Bacteria	C. albicans	E. coli	S. aureus	M. pneumonie
T Broth (48hrs / 37°C)	-	•	•	NA
Mycoview Screening				
Quantum (48hrs /	_	_	_	_
37°C)			_	_
(Reconstituted with T Broth)				

^{- :} No growth (no colour change and no turbidity appearance)

C. MYCOVIEW SCREENING QUANTUM FERTILITY AND TRAY CONTROL

Ureaplasma urealyticum

Sample	Results	Wells					
	(24 H)	С	Uu	Mh	L	Е	
Uu66 (10 ⁴ -10 ⁵ CCU/ml)	Expected	+	+	-	+	-	
	Obtained	+	+	-	+	-	
Uu66 (10 ² -10 ³ CCU/ml)	Expected	±	-	-	±	-	
	Obtained	±	-	-	±	-	
Uu S8 (10 ⁴ -10 ⁵ CCU/ml)	Expected	+	+	-	+	-	
	Obtained	+	+	-	+	-	

1

Uu S8 (10 ² -10 ³ CCU/ml)	Expected	±	•	-	±	1
	Obtained	±	ı	ı	±	ı
Uu CHB (10 ⁴ -10 ⁵ CCU/ml)	Expected	+	+	-	+	ı
	Obtained	+	+	-	+	
Uu CHB (10 ² -10 ³ CCU/ml)	Expected	±	-	-	±	-
	Obtained	±	-	-	±	-

Mycoplasma hominis

Sample	Results	Wells					
	(24-48 H)	С	Uu	Mh	L	E	
Mh PG21 (10 ⁴ -10 ⁵ CCU/ml)	Expected	+	-	+	-	+	
	Obtained	+	-	+	-	+	
Mh PG21 (10 ² -10 ³ CCU/ml)	Expected	±	-	-	-	±	
	Obtained	±	-	-	-	±	
Mh CHB (10 ⁴ -10 ⁵ CCU/ml)	Expected	+	-	+	1	+	
	Obtained	+	-	+	-	+	
Mh CHB (10 ² -10 ³ CCU/ml)	Expected	±	-	-	-	±	
	Obtained	±	-	-	-	±	

D. CONCLUSION

We herewith certify that the product has been checked in the Quality Control Department of ZEAKON Diagnostics SARL and has met the established Quality Control acceptance criteria.

Approved by : Pierre-Olivier WEICHLEIN Date:08/03/2018