

QUALITY CONTROL CERTIFICATE

PRODUCT NAME: MYCOVIEW SCREENING QUANTUM **CODE:** 2060Q

LOT NUMBER: 20211001A2SC

EXPIRY DATE: 29/04/2023

	LOT N°	EXPIRY DATE (dd/mm/yyyy)
T Broth	20210824	23/05/2023
Mycoview Screening Quantum	20210730	29/04/2023
Rayon Swabs	MFG: 20210801	31/07/2023

A. PHYSICAL INSPECTION

KIT COMPONENT	CHECKED
T Broth	<input checked="" type="checkbox"/>
Mycoview Screening Quantum	<input checked="" type="checkbox"/>
Rayon Swabs	<input checked="" type="checkbox"/>

B. INHIBITION TESTS

MEDIUM / Bacteria	C. albicans	E. coli	S. aureus	M. pneumonie
T Broth (48hrs / 37°C)	-	-	-	NA
Mycoview Screening Quantum (48hrs / 37°C) (Reconstituted with T Broth)	-	-	-	-

- : No growth (no colour change and no turbidity appearance)

C. MYCOVIEW SCREENING QUANTUM FERTILITY AND TRAY CONTROL

Ureaplasma urealyticum

Sample	Results (24 H)	Wells				
		C	Uu	Mh	L	E
Uu66 (10 ⁴ -10 ⁵ CCU/ml)	Expected	+	+	-	+	-
	Obtained	+	+	-	+	-
Uu66 (10 ² -10 ³ CCU/ml)	Expected	±	-	-	±	-
	Obtained	±	-	-	±	-
Uu S8 (10 ⁴ -10 ⁵ CCU/ml)	Expected	+	+	-	+	-
	Obtained	+	+	-	+	-

Uu S8 (10 ² -10 ³ CCU/ml)	Expected	±	-	-	±	-
	Obtained	±	-	-	±	-
Uu CHB (10 ⁴ -10 ⁵ CCU/ml)	Expected	+	+	-	+	-
	Obtained	+	+	-	+	-
Uu CHB (10 ² -10 ³ CCU/ml)	Expected	±	-	-	±	-
	Obtained	±	-	-	±	-

Mycoplasma hominis

Sample	Results (24-48 H)	Wells				
		C	Uu	Mh	L	E
Mh PG21 (10 ⁴ -10 ⁵ CCU/ml)	Expected	+	-	+	-	+
	Obtained	+	-	+	-	+
Mh PG21 (10 ² -10 ³ CCU/ml)	Expected	±	-	-	-	±
	Obtained	±	-	-	-	±
Mh CHB (10 ⁴ -10 ⁵ CCU/ml)	Expected	+	-	+	-	+
	Obtained	+	-	+	-	+
Mh CHB (10 ² -10 ³ CCU/ml)	Expected	±	-	-	-	±
	Obtained	±	-	-	-	±

D. CONCLUSION

We herewith certify that the product has been checked in the Quality Control Department of ZEAKON Diagnostics SARL and has met the established Quality Control acceptance criteria.

Approved by :

Date :18/11/2021

Pierre-Olivier WEICHLEIN