


## CERTIFICATE OF ANALYSIS

<b>Product Name:</b> Microbank™ (With 2D Barcodes) – Mixed		<b>Lot Number:</b> C20035M
<b>Product Code:</b> PL.170C/M		<b>Expiration Date:</b> 2023-08-31
<b>Product Code</b>	<b>Quantity</b>	<b>Components</b>
PL.170C/M	80	Vials printed with unique 2D barcodes containing approximately 25 beads and 1 ml of cryopreservative
	1	Numbered storage container with a unique 2D barcode
<b>Additional Information</b>		
Storage Conditions:		15-30°C
Precautions:		Refer to Instructions for Use
<b>Quality Control Testing</b>		
<b>Sterility*</b>	<b>Criteria</b>	<b>Result</b>
	Conforms	PASS
<b>Visual Inspection</b>	Conforms	PASS
<b>Barcode Verification/ Barcode Sequencing</b>	Conforms	PASS
<b>Organism</b>	<b>Recovery of Organism</b>	<b>Result</b>
<i>Escherichia coli</i> ATCC 25922	Satisfactory	PASS
<i>Staphylococcus aureus</i> ATCC 25923	Satisfactory	PASS
<i>Klebsiella pneumoniae</i> ATCC 13883	Satisfactory	PASS
<i>Clostridium perfringens</i> ATCC 13124	Satisfactory	PASS
<i>Neisseria gonorrhoeae</i> ATCC 43069	Satisfactory	PASS
*This product has been sterilized by GAMMA irradiation in a validated sterilization cycle and meets the established validated specified dose.		
<b>Release</b>		
2021-12-22		
Date	Quality Control	